Certificate of Calibration

Object:

Precision Sound Level Meter Nor131

Supplier:

Norsonic AS

Type:

Nor131

Serial number:

→ CONFIDENTIAL

Client:

This instrument is tested and calibrated in accordance to the Norsonic production standard set for Nor131, ensuring that the instrument conforms to the following standards;

IEC 61672-1:2002 class 1

IEC 61260-1 class 1 Ed 1.0 2014-02

ANSI S1.4-1983 (R2001) with amd. S1.4A-1985 class 1

ANSI S1.43-1997 (R2002) class 1

ANSI S1.11-2004 class 1 DIN 45 657, Applicable parts

IEC 61094 part 4

Instrumentation used for calibration traceable to:

Electrical Parameters: MT, Norway Acoustical Parameters: PTB, Germany

Environmental Parameters: Justervesenet. Norway

FW version:

3.1.984 2015-03-27 09:58r

ld no.:

CONFIDENTIAL

Accessories:

Preamplifier type:

1207 S.r

Microphone type: 1228

Sino

→ CONFIDENTIAL

Comments:

None

Date of calibration:

Calibration interval recommended:

2017-04-07

2 years

The environmental parameters applicable to this calibration are kept well within limits ensuring negligible deviation on obtained measurement results.

Calibrated by:

Anders Amundsen

Sign.

M