

Certificate of Calibration

Certificate No.: 

Object: Precision Sound & Vibration Analyser Nor150

Supplier: Norsonic AS

Type: Nor150

Serial number:



CONFIDENTIAL

Client:

This instrument is tested and calibrated in accordance to the Norsonic production standard set for Nor150, ensuring that the instrument conforms to the following standards;

IEC 61672-1:2002 class 1
IEC 61260-1 class 1 Ed 1.0 2014-02
ANSI S1.4-1983 (R2001) with amd. S1.4A-1985 class 1
ANSI S1.43-1997 (R2002) class 1
ANSI S1.11-2004 class 1
DIN 45 657, Applicable parts
IEC 61094 part 4

Instrumentation used for calibration traceable to:

Electrical Parameters: MT, Norway
Acoustical Parameters: PTB, Germany
Environmental Parameters: Justervesenet, Norway

Adjustments: None

Comments: None

Date of calibration:

2017-03-14

Calibration interval recommended

2 years

The environmental parameters applicable to this calibration are kept well within limits ensuring negligible deviation on obtained measurement results.

Calibrated by:

Sign. 